

Sample Result Name	Type	Measure Date Time	Recalculation Date Time	Origin
9752 (Ino-212043)	UNKNOWN	Saturday, July 26, 2014	4:34:24 PM	Measured
RX-37, RX-49, RX-45, BX-164, R-62, FX-66				
Method Name	Method Version	Operator Name	Check Type	Check Status
Fe-30-MO		SARA SAE	None	Not Used
Grade Verification Name	Grade Verification Similarity	Correction Type	Type Corr Sample Name	Grade Search Name
		None		
Grade Search Similarity	Type	Status	Outlier Removal	
0	None	Not Used		
Sample Name	Grade			
9752 (Ino-212043)				
RX-37, RX-49, RX-45, BX-164, R-62, FX-66				

	C	Si	Mn	P	S	Cr	Mo	Ni	Al	Co	Cu	Nb	Ti	V	W	Pb	Sn	As	Zr
	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc	Conc
%	%	%	%	%	%	%	%	%	%	%	%	%	%	%	%	%	%	%	%
2	0.032	0.18	0.38	0.030	0.033	0.017	<0.003	0.037	0.054	<0.005	0.005	<0.005	<0.0010	0.003	0.051	0.014	0.010	<0.005	0.006
3	0.031	0.18	0.37	0.030	0.042	0.020	<0.003	0.036	0.066	<0.005	0.006	<0.005	<0.0010	<0.002	<0.040	0.019	0.009	0.006	0.006
6	0.033	0.19	0.39	0.036	0.039	0.017	<0.003	0.037	0.031	<0.005	0.005	<0.005	<0.0010	0.003	0.043	0.014	0.011	0.009	0.006
Rep	0.032	0.18	0.38	0.032	0.038	0.018	0.003	0.037	0.050	<0.005	0.005	<0.005	<0.0010	0.003	0.045	0.015	0.010	0.006	0.006
Min Cal	0.006	0.005	0.004	0.003	0.002	0.005	0.003	0.005	0.003	0.005	0.002	0.005	0.0010	0.002	0.040	0.010	0.002	0.005	0.003
Mean	0.032	0.18	0.38	0.032	0.038	0.018	<0.011	0.037	0.050	<0.004	0.005	<0.010	<0.003	0.003	<0.032	0.015	0.010	0.006	0.006
Max Cal	3.00	5.04	18.12	0.18	0.30	36.00	10.20	55.2	6.24	21.60	7.44	3.60	3.60	12.00	7.32	0.13	0.27	0.13	0.19
	Ca	B	Fe																
	Conc	Conc	Conc																
	%	%	%																
2	0.002	0.001	99.1																
3	0.003	0.002	99.1																
6	0.002	0.002	99.1																
Rep	0.002	0.002	99.1																
Min Cal	0.0005	0.0005																	
Mean	0.002	0.002	99.2																
Max Cal	0.016	0.025																	

OK in 108 grade
as per spec 26-615
Deleah

